



Laboratory chemicals and products-Clinical chemistry and hematology products-Fine & industrial chemicals

## Certificate of Analysis

**Product No** 8404  
 Methanol  
 'BAKER HPLC ANALYZED'  
 HPLC Isocratic Grade

**Lot No** 1006728004

**Release Date** 08-Mar-10

Test	Specification	Actual value
Assay	min. 99.8 %	100.0 %
Color (APHA)	max. 10	< 5
Residue after Evaporation (in ppm)	max. 5	0.8
Titration Acid (meq/g)	max. 0.0003	0.00030
Titration Base (meq/g)	max. 0.0002	0.00003
Water (H <sub>2</sub> O)	max. 0.05 %	0.003 %
<i>Ultraviolet Absorbance (1.00-cm path vs water):</i>		
at 225 nm	max. 0.30	0.18
at 240 nm	max. 0.10	0.05
at 265 nm	max. 0.01	0.01

Use Before Date: 2012-04

For Laboratory, Research or Manufacturing Use.  
 This document is automatically generated.  
 The content was authorized by J. Mittendorf, QA & RA Manager  
 Mallinckrodt Baker B.V.  
 Teugseweg 20  
 7418 AM Deventer  
 The Netherlands  
 Tel:+31 570687500  
 For questions on this cofa, please contact Customer service at [jtbaker.nl@covidien.com](mailto:jtbaker.nl@covidien.com)  
 The information in this document is property of Mallinckrodt Baker B.V.