



Laboratory chemicals and products-Clinical chemistry and hematology products-Fine & industrial chemicals

## Certificate of Analysis

**Product No** 7098  
Sodium Hydroxide  
0.1 mol/l  
'BAKER ANALYZED'  
**Lot No** 0827502018  
**Release Date** 10-Oct-08

---

Volumetric Solution, ready for use

Test	Specification	Actual value
Titer (mol/l)	0.0997 - 0.1003	0.1000
<i>Trace Impurities (in ppm)</i>		
Chloride (Cl)	max. 5	< 1.5
Heavy Metals (as Pb)	max. 1	< 1.0
Iron (Fe)	max. 0.5	< 0.15

Each lot of this product is standardized potentiometrically against an NIST Standard Reference Material, Potassium Hydrogen Phthalate.

Use Before Date: 2010-10

For Laboratory, Research or Manufacturing Use.  
This document is automatically generated.  
The content was authorized by J. Mittendorf, QA & RA Manager  
Mallinckrodt Baker B.V.  
Teugseweg 20  
7418 AM Deventer  
The Netherlands  
Tel: +31 570687500  
Email: [jtbaker.nl@emea.tycohealthcare.com](mailto:jtbaker.nl@emea.tycohealthcare.com)  
The information in this document is property of Mallinckrodt Baker B.V.