



Laboratory chemicals and products-Clinical chemistry and hematology products-Fine &amp; industrial chemicals

## Certificate of Analysis

**Product No** 4218  
Water  
'BAKER HPLC ANALYZED'  
HPLC Gradient Grade

**Lot No** 1009200013

**Release Date** 08-Apr-10

---

Test	Specification	Actual value
Particulate Matter	max. 0.0001 %	< 0.0001 %
Residue after Evaporation	max. 2 ppm	< 2 ppm
<i>Ultraviolet absorbance of largest eluted peak</i>		
at 220 nm	max. 0.005 au	< 0.005 au
at 254 nm	max. 0.001 au	< 0.001 au
<i>Physical Data(not specifications):</i>		
Polarity Index	9.0	
Solvent Group	9	

Use Before Date: 2012-04

For Laboratory, Research or Manufacturing Use.  
This document is automatically generated.  
The content was authorized by J. Mittendorf, QA & RA Manager  
Mallinckrodt Baker B.V.  
Teugseweg 20  
7418 AM Deventer  
The Netherlands  
Tel:+31 570687500  
For questions on this cofa, please contact Customer service at [jtbaker.nl@covidien.com](mailto:jtbaker.nl@covidien.com)  
The information in this document is property of Mallinckrodt Baker B.V.